

SOLUTION CHANGE

BFD

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 08/01/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:03
INTERNAL STANDARD	VERIFIED	22:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07

