

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 01/11/2022

OPERATOR NAME:  
LEONARD F DILLENBECK  
Permit #: 47621  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:03
INTERNAL STANDARD	VERIFIED	01:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:04
BLANK TEST	0.00	01:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:05
BLANK TEST	0.00	01:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:06
BLANK TEST	0.00	01:07

