

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 03/16/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:30
INTERNAL STANDARD	VERIFIED	17:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:30
BLANK TEST	0.00	17:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:31
BLANK TEST	0.00	17:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:32
BLANK TEST	0.00	17:33

