

BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/27/2022

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:44
INTERNAL STANDARD	VERIFIED	17:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:44
BLANK TEST	0.00	17:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:45
BLANK TEST	0.00	17:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:47
BLANK TEST	0.00	17:48

