

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 07/06/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:04
INTERNAL STANDARD	VERIFIED	03:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:04
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:06
BLANK TEST	0.00	03:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:07
BLANK TEST	0.00	03:08

