

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 10/05/2022

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:37
INTERNAL STANDARD	VERIFIED	17:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:38
BLANK TEST	0.00	17:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:39
BLANK TEST	0.00	17:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:40
BLANK TEST	0.00	17:41

