

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 07/02/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:59
INTERNAL STANDARD	VERIFIED	20:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:59
BLANK TEST	0.00	21:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:00
BLANK TEST	0.00	21:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:02
BLANK TEST	0.00	21:02

