

**SUPERVISOR TEST**

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 141906

*BFO*

Date: 05/10/2023

OPERATOR NAME:  
 STEVEN P EILERS  
 Permit #: 49321  
 Operator Agency/Dept: ONEIDA CO. S.O.  
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:33
INTERNAL STANDARD	VERIFIED	10:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:33
BLANK TEST	0.00	10:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:36
BLANK TEST	0.00	10:37

