

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 06/07/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:15
INTERNAL STANDARD	VERIFIED	23:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:16
BLANK TEST	0.00	23:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:17
BLANK TEST	0.00	23:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:18
BLANK TEST	0.00	23:19

