

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/14/2023

BFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:54
INTERNAL STANDARD	VERIFIED	00:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:54
BLANK TEST	0.00	00:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:56
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:58

