

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/29/2023

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:32
INTERNAL STANDARD	VERIFIED	05:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:33
BLANK TEST	0.00	05:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:34
BLANK TEST	0.00	05:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:35
BLANK TEST	0.00	05:36

