

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

*BFO*

Date: 08/25/2023

OPERATOR NAME:  
KRISTY N ALLEN  
Permit #: 49726  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:11
INTERNAL STANDARD	VERIFIED	00:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:13
BLANK TEST	0.00	00:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:14
BLANK TEST	0.00	00:15

