

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 09/14/2023

OPERATOR NAME:  
ARTUR KIRILKO  
Permit #: 52001  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:12
INTERNAL STANDARD	VERIFIED	00:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:14
BLANK TEST	0.00	00:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:15
BLANK TEST	0.00	00:16

