

SUPERVISOR TEST

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 11/29/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:47
INTERNAL STANDARD	VERIFIED	22:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:47
BLANK TEST	0.00	22:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:48
BLANK TEST	0.00	22:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:50
BLANK TEST	0.00	22:50

