

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 12/27/2023

OPERATOR NAME: JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:54
INTERNAL STANDARD	33.99c	22:54
REFERENCE STANDARD TEMP	0.09	22:54
BLANK TEST	0.00	22:55
REFERENCE STANDARD TEMP	33.99c	22:55
REFERENCE STANDARD	0.09	22:55
BLANK TEST	0.00	22:56
REFERENCE STANDARD TEMP	33.99c	22:56
REFERENCE STANDARD	0.10	22:57
BLANK TEST	0.00	22:57

