

SUPERVISOR TEST

BFD

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 02/14/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:22
INTERNAL STANDARD	VERIFIED	03:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:22
BLANK TEST	0.00	03:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:24
BLANK TEST	0.00	03:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:25
BLANK TEST	0.00	03:26

