

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/29/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:49
INTERNAL STANDARD	VERIFIED	22:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:49
BLANK TEST	0.00	22:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:51
BLANK TEST	0.00	22:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53

