

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 12/08/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:47
INTERNAL STANDARD	VERIFIED	21:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:48
BLANK TEST	0.00	21:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:49
BLANK TEST	0.00	21:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:50
BLANK TEST	0.00	21:51

