

SOLUTION CHANGE

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 06/01/2022

CFO

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:22
INTERNAL STANDARD	VERIFIED	02:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:23
BLANK TEST	0.00	02:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:24
BLANK TEST	0.00	02:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:25
BLANK TEST	0.00	02:26

