

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 02/16/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:11
INTERNAL STANDARD	VERIFIED	23:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:12
BLANK TEST	0.00	23:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:13
BLANK TEST	0.00	23:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:14
BLANK TEST	0.00	23:15

