

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

CFO

Date: 04/20/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:19
INTERNAL STANDARD	VERIFIED	05:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:20
BLANK TEST	0.00	05:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:21
BLANK TEST	0.00	05:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:22
BLANK TEST	0.00	05:23

