

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/08/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:19
INTERNAL STANDARD	VERIFIED	00:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:20
BLANK TEST	0.00	00:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:21
BLANK TEST	0.00	00:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:22
BLANK TEST	0.00	00:23

