

SUPERVISOR TEST

 ONEIDA CO.
 SHERIFF'S DEPT
 DATAMASTER dmt: 123906

Date: 07/27/2022

OPERATOR NAME:
 NICHOLAS A MARCANO
 Permit #: 49335
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:14
INTERNAL STANDARD	VERIFIED	06:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:15
BLANK TEST	0.00	06:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:16
BLANK TEST	0.00	06:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:17
BLANK TEST	0.00	06:18

CFO


