

SUPERVISOR TEST

 ONEIDA CO.
 SHERIFF'S DEPT
 DATAMASTER dmt: 123906

Date: 09/01/2022

OPERATOR NAME:
 STEVEN P EILERS
 Permit #: 49321
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:51
INTERNAL STANDARD	VERIFIED	00:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:53
BLANK TEST	0.00	00:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:54
BLANK TEST	0.00	00:55

