

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 09/07/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:26
INTERNAL STANDARD	VERIFIED	02:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:27
BLANK TEST	0.00	02:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:29
BLANK TEST	0.00	02:29

