

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 09/14/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:39
INTERNAL STANDARD	VERIFIED	22:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:39
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:41
BLANK TEST	0.00	22:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:42
BLANK TEST	0.00	22:43

