

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/28/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:39
INTERNAL STANDARD	VERIFIED	02:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:42
BLANK TEST	0.00	02:43

