

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/09/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:58
INTERNAL STANDARD	VERIFIED	04:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:58
BLANK TEST	0.00	04:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:59
BLANK TEST	0.00	05:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:00
BLANK TEST	0.00	05:01

|| |

