

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 12/07/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:23
INTERNAL STANDARD	VERIFIED	02:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:24
BLANK TEST	0.00	02:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:26
BLANK TEST	0.00	02:27

CFO

