

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 12/21/2022

OPERATOR NAME:  
NOAH A DANQUER  
Permit #: 50812  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:23
INTERNAL STANDARD	VERIFIED	01:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:23
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:24
BLANK TEST	0.00	01:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:26
BLANK TEST	0.00	01:27

CFO

