

SOLUTION CHANGE

CFD

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 09/01/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:21
INTERNAL STANDARD	VERIFIED	23:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:21
BLANK TEST	0.00	23:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:23
BLANK TEST	0.00	23:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:24
BLANK TEST	0.00	23:25

