

SOLUTION CHANGE

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 03/28/2023

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:47
INTERNAL STANDARD	VERIFIED	21:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:47
BLANK TEST	0.00	21:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:49
BLANK TEST	0.00	21:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:50
BLANK TEST	0.00	21:51

