

SOLUTION CHANGE

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 12/01/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:28
INTERNAL STANDARD	VERIFIED	01:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:28
BLANK TEST	0.00	01:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:30
BLANK TEST	0.00	01:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:31
BLANK TEST	0.00	01:32

