

SUPERVISOR TEST

CFD

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 06/29/2023

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:50
INTERNAL STANDARD	VERIFIED	05:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:51
BLANK TEST	0.00	05:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:52
BLANK TEST	0.00	05:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:53
BLANK TEST	0.00	05:54

