

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/24/2023

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:41
INTERNAL STANDARD	VERIFIED	19:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:42
BLANK TEST	0.00	19:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:43
BLANK TEST	0.00	19:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:44
BLANK TEST	0.00	19:45

