

SUPERVISOR TEST

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 10/25/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:11
INTERNAL STANDARD	VERIFIED	01:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:11
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:15

