

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 02/22/2023

OPERATOR NAME: CFO
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:49
INTERNAL STANDARD	VERIFIED	22:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:49
BLANK TEST	0.00	22:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:51
BLANK TEST	0.00	22:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53

