

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 03/10/2023

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:45
INTERNAL STANDARD	VERIFIED	04:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:46
BLANK TEST	0.00	04:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:47
BLANK TEST	0.00	04:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:48
BLANK TEST	0.00	04:49

