

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 11/15/2023

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:42
INTERNAL STANDARD	VERIFIED	17:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:42
BLANK TEST	0.00	17:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:43
BLANK TEST	0.00	17:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:45
BLANK TEST	0.00	17:46

