

Infrared Instrument

Weekly Calibration Verification Test Record

Department: OC50

Instrument Serial #: 123906

| Test Date | Simulator Solution Lot # | Simulator Temperature | | | Sim. Serial# | Test Results | | | Breath Analysis Supervisor | Sim. Sol. Change * |
|-----------|--------------------------|-----------------------|-------|-------|--------------|--------------|-----|-----|----------------------------|--------------------|
| | | 1 | 2 | 3 | | 1 | 2 | 3 | | |
| 9/27/23 | 23190 | 34.01 | 34.01 | 34.01 | MP4189 | .10 | .10 | .10 | Eilers | No |
| 10/01 | 23270 | 34.00 | 34.01 | 34.01 | MP4189 | .10 | .10 | .10 | Cusack | Yes |
| 10/01 | 23270 | 34.01 | 34.01 | 34.01 | MP4189 | .10 | .10 | .10 | Dixon | NO |
| 10/11/23 | 23270 | 34.01 | 34.01 | 34.01 | MP4189 | .09 | .09 | .10 | Osika | NO |
| 10/18/23 | 23270 | 34.01 | 34.01 | 34.01 | MP4189 | .10 | .09 | .09 | Cusack | No |
| 10/25/23 | 23270 | 34.01 | 34.01 | 34.01 | MP4189 | .09 | .09 | .09 | Osika | NO |
| 11/1/23 | 23270 | 34.01 | 34.00 | 34.00 | MP4189 | .10 | .10 | .10 | Dixon | Yes |
| 11/8/23 | 23270 | 34.00 | 34.01 | 34.00 | MP4185 | .10 | .10 | .10 | Osika | No |
| 11/15/23 | 23270 | 34.01 | 34.01 | 34.01 | MP4189 | .10 | .09 | .10 | Dixon | NO |
| 11/21/23 | 23270 | 34.01 | 34.01 | 34.01 | MP4189 | .09 | .09 | .09 | Dixon | NO |
| 11/29/23 | 23270 | 34.01 | 34.01 | 34.01 | MP4189 | .09 | .09 | .09 | Osika | NO |
| 12/1/23 | 23350 | 34.01 | 34.01 | 34.01 | MP4189 | .10 | .10 | .10 | Osika | Yes |

* - Please enter a √ to denote simulator solution change taking place

CFO