

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 02/08/2024

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:19
INTERNAL STANDARD	VERIFIED	04:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:20
BLANK TEST	0.00	04:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:21
BLANK TEST	0.00	04:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:22
BLANK TEST	0.00	04:23

