

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 01/10/2024

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:52
INTERNAL STANDARD	VERIFIED	17:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:52
BLANK TEST	0.00	17:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:53
BLANK TEST	0.00	17:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:55
BLANK TEST	0.00	17:55

