

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/01/2021

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:54
INTERNAL STANDARD	VERIFIED	07:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:55
BLANK TEST	0.00	07:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:56
BLANK TEST	0.00	07:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:57
BLANK TEST	0.00	07:58

