

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 09/22/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:58
INTERNAL STANDARD	VERIFIED	23:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:59
BLANK TEST	0.00	23:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:00
BLANK TEST	0.00	00:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:01
BLANK TEST	0.00	00:02

