

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 10/13/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:56
INTERNAL STANDARD	VERIFIED	21:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:57
BLANK TEST	0.00	21:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:58
BLANK TEST	0.00	21:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:59
BLANK TEST	0.00	22:00

