

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/27/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:19
INTERNAL STANDARD	VERIFIED	20:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:19
BLANK TEST	0.00	20:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:20
BLANK TEST	0.00	20:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:22
BLANK TEST	0.00	20:23

