

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 12/08/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:13
INTERNAL STANDARD	VERIFIED	09:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	09:13
BLANK TEST	0.00	09:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	09:14
BLANK TEST	0.00	09:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	09:15
BLANK TEST	0.00	09:16

