

SOLUTION CHANGE

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/01/2022

OPERATOR NAME:
AARON M ALSHAMAN
Permit #: 50807
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:13
INTERNAL STANDARD	VERIFIED	19:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:14
BLANK TEST	0.00	19:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:15
BLANK TEST	0.00	19:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:16
BLANK TEST	0.00	19:17

