

SUPERVISOR TEST

W/S

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 01/26/2022

OPERATOR NAME:
ZACHARY S PRZYBYLA
Permit #: 49339
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:53
INTERNAL STANDARD	VERIFIED	22:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:53
BLANK TEST	0.00	22:54
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:55
BLANK TEST	0.00	22:55
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:56
BLANK TEST	0.00	22:57

