

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

*WFO*

Date: 02/16/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:52
INTERNAL STANDARD	VERIFIED	17:52
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:52
BLANK TEST	0.00	17:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:53
BLANK TEST	0.00	17:54
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:54
BLANK TEST	0.00	17:55

